



CC-2857

TESLA/FF-050

Page :1 of 2

Calibration Certificate

Certificate No	LM250521/2 / TC / FL / 002	Unique Lab Report (ULR) ID	CC23452500000129F
Date of Calibration	21/05/2025	Recommended Next Due Date*	21/05/2026
Receipt / WRF No.	LM250521/2	Date of DUC Received	21/05/2025
Certificate Issue Date	22/05/2025	Calibration Performed at	Lab

Details of Customer

Name	AIPL
Address	B-95, Sector-25 GIDC, Gandhinagar, Gujarat, India, 382016
Ref. Doc. No. (GP/PO)	Gp_001
Date of Doc.	5/21/2025

Details of Device under Calibration (DUC)

Nomenclature	Digital Flow meter (Electromagnetic)		
Range of Meter	0.2 to 9 m ³ /hr	Mfg. Serial No	EFM/2025/109
Resolution	0.01 m ³ /hr	Identification No	109
Size	25 NB (1")	Model No	AIPL-EFM-50NB-PTFE
Make	Accumax	Condition of DUC	Good

Details of Calibration Masters and Traceability

Lab ID	Description of Master	Traceability
TEQ030	Mass Flow meter, Make: Krohne, Model: IFS4000, Size: DN25, Serial No. G140000008305187, Range: 0 to 5 m ³ /hr	Traceable to National Standard through NABL Lab Certif. No CC-3526, Ahmedabad, Certificate No. IPC-11092024-047, Cal Date : 11/09/2024, Valid up to 10/09/2025

Details of Calibration Procedure and Reference Documents

Lab Procedure	Method of Calibration
Lab Procedure No. : COP/005-01	The calibration of DUC was carried out by comparison with Standard flow meter by using Calibration Liquid as Water
Calibration Method is generally based	NABL 129-2019

Details of Environment Condition at the Time of Calibration

Temperature	29.0 °C	Relative Humidity	34 % Rh
Temperature of Calibration Liquid	22.50 °C	Density of Calibration Liquid	1.0 ± 0.002 Sp. Gr.
Uncertainty of Measurement		refer next page	

"The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$ such that the coverage probability corresponds to 95.45% confidence level



To verify the basic details of the Calibration Certificate kindly scan the QR code and follow the link provided.



CC-2857

Signature Not Verified

Digitally Signed By:
DS TESTING ELECTRICAL STANDARDS LABORATORY AND
R&D INSTITUTE PRIVATE LIMITED 1
Wed 31-Dec-2025 15:38:42 IST
Approved by Vinod Kotadiya (Tech. Manager)

Calibration Certificate

Certificate No	LM250521/2 / TC / FL / 002	Unique Lab Report (ULR) ID	CC234525000000129F
Date of Calibration	21/05/2025	Recommended Next Due Date*	21/05/2026

DISCIPLINE : FLUID FLOW CALIBRATION

GROUP : FLOW MEASURING DEVICES

Calibration Result

1.0 Calibration Result of DUC / Error of Measurement

Calibration is performed by considering the SI unit of Flow $1 \text{ m}^3/\text{hr} = 1 \text{ m}^3/\text{hr}$

Sr. No.	Reading on Master (m^3/hr)	Reading on DUC (m^3/hr)	Error (m^3/hr)	Error in % of Reading (%)	Expanded Uncertainty (%)	Coverage Factor
1	0.5131	0.56	0.0469	9.141	± 1.64	2.00
2	2.5234	2.55	0.0266	1.054	± 1.64	2.00
3	3.9908	4.00	0.0092	0.231	± 1.64	2.00
4	5.9908	6.00	0.0092	0.154	± 1.64	2.00
5	8.4908	8.50	0.0092	0.108	± 1.64	2.00

→ K Factor (Primary Constant) of Flow meter at the time of Calibration : 0.23335

General Remarks

- Reporting of DUC/Master Reading is an average of five readings
- The reading has been rounded off wherever applicable
- Calibration is performed without doing any adjustment in its original condition at time of receipt/calibration, unless mentioned herein
- The calibration results relate only to the item calibrated and result reported in the certificate are valid at the time of and under the stated conditions of measurement.
- The Calibration certificate shall not be reproduced except in full, without written permission of the Laboratory
- *Recommended Next Due Date is reported as per requisition of customer

-- End of Certificate --



To verify the basic details of the Calibration Certificate kindly scan the QR code and follow the link provided.



CC-2857

Signature Not Verified

Digitally Signed By:
DS TESTING ELECTRICAL STANDARDS LABORATORY AND
R&D INSTITUTE PRIVATE LIMITED 1
Wed 31-Dec-2025 15:38:43 IST
Approved by Vinod Kotadiya (Tech. Manager)